

Ohlro Medisafe



Validation using 6 ThermaBrick® Refrigerant Bricks

TEMPRECISION
INTERNATIONAL

Test configuration MediSafe L

- ISTA Summer profile
- Payload 7x250g sample bottles filled with water
- Samples and ThermaBrick® Refrigerant Bricks conditioned at 3° C



Results

- This configuration lasted over 120h
- Fail criteria: $< 2^{\circ}\text{C}$ or $> 8^{\circ}\text{C}$

